

Beveled Cannula Placement for 23 and 25 Gauge Vitrectomy

Robert L. Avery, MD (Santa Barbara, CA)*

SYNOPSIS

Beveled entry with 23 and 25 gauge trocar/cannula systems for vitrectomy are demonstrated. The entry is rapid and does not require specialized instruments. This simple modification of the introduction of the cannula allows for a self-sealing wound – thereby potentially lowering the risk of hypotony, wound leak, and endophthalmitis.